

ACCUNIQ BC380

The New Standard in Body Composition Analysis

BODY COMPOSITION ANALYZER

Multi-Frequency Segmental Body Composition Analysis using BIA Technology



02 Product Introduction ACCUNIQ BC380

Accurate Analysis and Seamless Data Management

ACCUNIQ BC380 delivers clinically accurate body composition results in less than 1 minute.

Easy to use, with on-screen step by step instructions and loaded with all the measurement data you will need to fully assess your clients composition levels.

On Screen results and full page report shows measurements with healthy ranges for quick assessment.






correlation with DEXA equipment (Lean body mass R^2 =0.9532), analyzes body composition and monitors body fat and muscle changes.



Touch panel and intuitive UI

The 7-inch wide color touchscreen and convenient keypad combined with on-screen instructions makes the BC380 easy for anyone to operate.





| ACCUNIQ | 6234 | Height 150 | 5.0 | Age 28 | 3 | Gender Fem | ale |
|---------------------|--------------|------------------------------|----------|----------|---------------------|---------------------|-----|
| | 0 × 0 × 0 × | Measuremen | Results | | | | |
| Comprehensive | Evaluation — | Body Compo | sition A | Inalysis | | | |
| Body Type | Standard | Heasure | | | | | |
| | | TBW 26.7 L (27.4-31.2) | 26.7 | | | | |
| Biological Age | 29 years | Proteins 7,3 kg (7.5-6.6) | | 34.0 | | | > |
| Basal Metabolic Rat | te 1159 kcal | Minerals 2.5 kg (2.9-3.2) | | | 36.5 (37.5-42.0) | | |
| Body Cell Mass | 24.9 kg | MBF 14.0 kg (10.7-16.6) | | | | 50.5 (45.5-61.6) | |
| Summary | Details | History | F | rint | | Exit | |



Electrode handle

The measurement starts automatically without the need to push a button separately when you grab the electrode handle that gives you a sense of improved grip.

Convenient foothold electrode

large 18"x18" platform and low profile allows for easy access, comfort and stability.

Enhanced expandability by connecting the product to various devices



Thermal printer
Thermal printer for fast printed results.



Ankle electrodes

No need to take off your socks.



Ultrasonic Height Meter

The ultrasonic height meter increases precision and automatically inputs height.



Fully automatic Blood Pressure Monitor

You can diagnose obesity and measure blood pressure simultaneously by connecting a blood pressure monitor to ACCUNIQ BC380.

04 Management ACCUNIQ BC380

Streamline Measurement Data Analysis






Body composition management program $^{\mathsf{ACCUNIQ}}$ $^{\mathsf{MANAGER}}$

ACCUNIQ MANAGER is a client health data management solution that captures and presents all body composition results with comparisons to healthy ranges for fast and easy assessment.

- · Generate historical comparisons to show progress and helps you personalize your diet and exercise prescriptions based on critical data.
- · Print full page reports, thermal receipt, or transfer results to devices.
- · Recommendation of personalized diet and exercise program to achieve healthy range.
- Easy-to-understand, Intuitive screen layout with graphical data presentation.
- · Screen showing systematic body composition analysis results for individual users and the provides for efficient data management processing.



Body composition management application ACCUNIQ APP

ACCUNIQ App is a mobile service for the management of personal body composition data. You can scan the QR code of the body composition measurement result with a smart phone so clients can check the result anytime.

- · Shows graphical statistical views of composition results and changes.
- · A Control Guide shows recommendations on caloric intake and excercise levels to reach healthy ranges.
- · Stores all measurements and shows historical graphing of measurement levels and healthy ranges.



Remote support program ACCUNIQ REMOTE SUPPORT

You can use the USB Wi-Fi dongle, which is provided for easy use of the device, to connect to the Selvas Healthcare C/S division and receive remote service more conveniently when checking or repairing the device.

- · Even if you don't know how to use the device, you can set its options through remote control.
- · You can remove measurement errors by checking the offset value, performing calibration, etc.
- · You can recover the firmware remotely from a PC.

06 Result Sheet ACCUNIO BC380

Comprehensive Results Page

Results are printed in easy to understand format with measurement comparison to the healthy range, making analysis fast and concise.



Body Composition Analysis

Shows the measurement results and normal range of total body water, protein, minerals and body fat which equal total body weight.

Muscle / Fat Analysis

Graph shows results for weight, skeletal muscle mass, and body fat mass compared to normal range.

Obesity Analysis

Graph shows body mass index and body fat percentage, compared to healthy range important indicators of obesity.

4 Abdominal Obesity Analysis

The fat of the human body consists of subcutaneous fat and visceral fat. This analysis assesses visceral fat that are closely related to adult diseases by using various indicators.

5 Segmental Lean and Fat Analysis

Graph shows muscle mass and fat mass of each of the five body parts (left arm, right arm, left leg, right leg, and torso).

(b) Body Composition Change

Historical graph of weight, skeletal muscle mass, and body fat mass, important indicators to assess progress.

Comprehensive Evaluation

Shows body type, body age, basal metabolic rate, calories needed per day, body cell mass, visceral fat mass, degree of obesity.

Body Balance Assessment

Assessement of the balance between the left and right of the body, and the upper and lower parts of the body. It evaluates whether the body maintains balance between the left and right of the body and the upper and lower parts of the body, rather than about the mass of skeletal muscle or fat.

Control Guide

Extracellular water ratio indicates the ratio of extracellular body water to total body water. This index evaluates the body's water balance and displays the body's current state as normal, boundary, or abnormal.

Segmental Lean Mass


Muscle mass and status of the five body parts (left arm, right arm, left leg, right leg, torso) compared to the standard weight.

Impedance

Indicates impedance by frequencies and by body parts. Impedance is a resistance generated when electric current passes through the body. Each person has a unique impedance.

Blood Pressure Analysis

Shows the blood pressure data when the device is connected to the hematomanometer provided by ACCUNIQ. This is especially useful because it assesses your obesity level and blood pressure at the same time.



Results Page for Child and Youth (Optional)

8

9

1

0

1

ACCUNIQ BC380

ID/NAME: 3456/SELVAS

Height: 187.5 cm Age: 47 years Gender: Male

Test Date/Time: 6/14/2017 15:34

Body Composition Analysis

| | values | Body Water | Soft Lean Mass | Fat-Free Mass | Weight |
|------------|-----------------------|------------|-----------------------|---------------|---------------|
| Body Water | 48.4 (45.1 ~ 48.0) | 48.4 | 00.0 | | |
| Proteins | 13.6 (12.4 ~ 13.9) | | 62.0 (53.3 ~ 65.1) | 66.2 | |
| Minerals | 4.3 (4.5 ~ 4.6) | | | (51.8 ~ 65.7) | 85.2 |
| Body Fat | 19.0 | | | | (65.7 ~ 88.9) |

— Muscle/Fat Analysis

2

3

4

6

6

| | U | nder | | Norma | al 📗 | | | | Ove | ſ | | | |
|-----------------------------|----|------|----|-------|----------------------------------|-------------|-----|-----|-----|-----|-----|-----|-----|
| Weight | 65 | 75 | 85 | 100 | 1 ¹ 5 8 5.2 | 125 | 135 | 145 | 155 | 165 | 175 | 185 | [%] |
| SMM Skeletal Muscle Mass | 70 | 80 | 90 | 100 | 110 37.2 | 120 | 130 | 140 | 150 | 160 | 170 | 180 | [%] |
| Fat Mass | 40 | 60 | 80 | 100 | 120 | 170 19.0 | 220 | 270 | 320 | 370 | 420 | 470 | [%] |

Obesity Analysis

| | Un | der | | Norma | ı | | | | Over | | | | |
|--------------------------------------|-------|-------|-------|-------|-------------|-------|-------|-------|-------|-------|-------|---------|---------|
| BMI | 14.50 | 16.50 | 18.50 | 21.75 | 25.00 | 27.21 | 29.42 | 31.64 | 33.85 | 36.07 | 38.28 | 40.50 [| [kg/m²] |
| Body Mass Index kg/m ² | | | | | = 24 | .2 | | | | | | | |
| PBF | 10.0 | 12.5 | 15.0 | 17.5 | 20.0 | 26.4 | 32.8 | 39.2 | 45.7 | 52.1 | 58.5 | 65.0 | [%] |
| Percentage of Body Fat | | | | | | 22.3 | | | | | | | |

Abdominal Obesity Analysis

| | Under | Normal | | Over | |
|--------------------|--------------|----------|----------|--------------------|---------------------|
| WHR | 0. | 75 O. | 90 | | |
| Waist to Hip Ratio | | | 0.90 | | |
| | Subcutaneous | Balanced | Boundary | Visceral Obesity I | Visceral Obesity II |
| VFL | 1 | 5 | 9 '11 | | 16 |
| Visceral Fat Level | | | | 1 2 | |
| VFA | 1 | 50 1 | бо | | |
| Visceral Fat Area | | | 114 | | |

Segmental Lean and Fat Analysis



Body Composition Change

| Weight | 81.8 | 83.5 | 85.2 | 85.2 | | |
|------------------------------------|-----------------------|-----------------------|-----------------------|-----------------------|------|------|
| SMM Skeletal Muscle Mass | 37.4 | 37.8 | 37.1 | 37.2 | | |
| PBF Percentage of Body Fat % | 18.8 | 19.6 | 22.5 | 22.3 | | |
| Test date | 2017.04.17 (14:00) | 2017.05.13 (12:10) | 2017.06.12 (15:46) | 2017.06.14 (15:34) | | |



Comprehensive Evaluation

| Body Type | over fat | class 1 | |
|----------------------|------------|---------|-------|
| Biological Age | | 47 | years |
| Basal Metabolic Ra | te(BMR) | 1800 | kcal |
| Total Daily Energy E | xpenditure | 2772 | kcal |
| Body Cell Mass | | 45.6 | |
| | | 0 = | |

Visceral Fat Mass 2.7

Body Balance Assessment -

| Upper Body L/R | balanced imbalanced I | $\begin{tabular}{ c c c c c c }\hline & imbalanced & \mathbb{I}\\\hline \end{tabular}$ |
|----------------|-----------------------|--|
| Lower Body L/R | balanced imbalanced I | $\ \ \ \ \ \ \ \ \ \ \ \ \ $ |

Control Guide

| Target Weight | 79.7 | kg |
|----------------|------|----|
| Weight Control | -5.5 | kg |
| Muscle Control | +0.0 | kg |
| Fat Control | -5.5 | kg |

ECW ratio 0.386 (Optimal)

Segmental Lean Mass (Based on standard weight)

| Right Arm | 3.72 kg | [3.12 ~ 3.81] | / Fit |
|-----------|----------|-----------------|-------|
| Left Arm | 3.60 kg | [3.12 ~ 3.81] | / Fit |
| Trunk | 28.02 kg | [23.46 ~ 28.67] | / Fit |
| Right Leg | 10.28 kg | [8.61 ~ 10.52] | / Fit |
| Left Leg | 10 17 kg | [861 ~ 1052] | / Fit |

Impedance

| Freq | 5K | 50K | 250K |
|---------|-----|-----|------|
| RA,Imp. | 275 | 243 | 207 |
| LA,Imp. | 281 | 254 | 214 |
| Trunk | 36 | 33 | 28 |
| RL,Imp. | 185 | 163 | 139 |
| LL,Imp, | 189 | 174 | 144 |

Blood Pressure Analysis



For history management please upload this results at the website using QR code scanning.

 $JUN\ 15,\,2017\ Rev\ A.1\ \ \mbox{SELVAS Healthcare, Inc.}$

ACCUNIQ BC380 Specifications

| Model | ACCUNIQ BC380 |
|---------------------------------|--|
| Measurement Method | Tetra-polar electrode method using 8 touch electrodes |
| Frequency Range | 5, 50, 250 kHz |
| Measurement Area | Whole body and Segmental measurement (arms, legs, and trunk) |
| Result Sheet Data | [Result for Body Composition Analysis] Body Composition Analysis (Weight, Lean Body Mass, Body Fat Mass, Muscle Mass, Protein Mass, Mineral Mass, Total Body Water), Skeletal Muscle / Fat Analysis (Weight, Skeletal Muscle Mass, Body Fat Mass), Obesity Analysis (Body Mass Index, Body Fat Percentage, Degree of Obesity, Abdominal Circumference), Abdominal Obesity Analysis (Waist Hip Ratio, Visceral Fat Level, Visceral Fat Area, Visceral Fat Mass), Cumulative Body Change Graph (Weight, Skeletal Muscle Mass, Body Fat Percentage), Overall Evaluation (Body Type, Body Age, Basal Metabolic Rate, Calories Needed per Day, Body Cell Mass, Total Score), Body Balance Assessment (Left and Right of the Upper Body, Left and Right of the Lower Body), Weight Control Targets (Recommended Weight, Weight Control Value, Muscle Control Value, Fat Control Value), Extracellular Water Ratio, Body Fat Mass / Muscle Mass by Body Parts (Left Arm, Right Arm, Left Leg, Right Leg, Torso), Impedance (By Body Parts and By Frequencies), Blood Pressure (When Interlocked with the Hematomanometer), QR Code [Result for Child and Youth (optional)] Body Composition Analysis (Weight, Lean Body Mass, Body Fat Mass, Muscle Mass, Protein Mass, Mineral Mass, Total Body Water), Skeletal Muscle / Fat Analysis (Weight, Skeletal Muscle Mass, Body Fat Mass), Obesity Analysis (Body Mass Index, Body Fat Percentage, Waist Hip Ratio), Child Growth Curve (Height, Weight), Nutritional Assessment (Protein, Mineral, Fat), Overall Evaluation (Body Type, Basal Metabolic Rate, Calories Needed per Day, Body Cell Mass, Degree of Obesity), Body Balance Assessment (Left and Right of the Upper Body, Left and Right of the Lower Body), Weight Control Targets (Recommended Weight, Weight Control Value, Muscle Control Value, Fat Control Value), Body Fat Mass / Muscle Mass by Body Parts (Left Arm, Right Arm, Left Leg, Right Leg, Torso) Impedance (By Body Parts and By Frequencies), QR Code |
| Measurement Current | Approx. 180 μA ± 15 |
| Power Consumption | 60 VA |
| Power Supply Voltage | Input: 100-240 VAC, 50/60 Hz, 1.4-0.7 A Output: DC 12 V, 5.0 A, 60 W MAX |
| Display | 7-inch Wide Color LCD |
| Input Device | Touchpad, Keypad |
| Transmission Device | 5 USB ports, 2 RS-232C ports, Wi-Fi (basic), Bluetooth (optional) |
| Printing Device | USB port (printer designated by the manufacturer), thermal printer (optional) |
| Dimensions | Main Unit 641×436×1029 mm (WxD×H±10 mm) Main Unit+Height Meter 795×436×2327 mm (WxD×H±10 mm) |
| Weight | Approx. 18kg (main unit) |
| Measurement Range | 100 - 950 Ω |
| Measurement Time | Approx. 30 sec. |
| Applicable Height | 50 - 220 cm |
| Measurement Height | 100 - 210 cm |
| Measurement Weight | 10 - 250 kg |
| Applicable Age | 1- 99 years old |
| Operating Environment | Temperature 5 - 40°C, relative humidity 15 - 93% (no condensation) |
| Storage Environment | Temperature -25 - 70°C, relative humidity less than 93% (no condensation) |
| Optional Equipment | Ultrasonic anthropometer, fully automatic hematomanometer, ankle electrodes, result sheet for Infants, USB memory, thermal printer, and Bluetooth |
| Printing Logo | Printing the hospital name, address, contact information, and logo available |
| Touch Screen | Touch screen's sensor location adjustable |
| Data Storage | Up to 100,000 data units can be stored when using an ID. |
| Measurement Mode | Scale mode / Body composition mode |
| Various Result Sheets | Body composition result sheet, Result sheet for Infants (Optional) |
| Checking Measurement Results | LCD, Internet, ACCUNIQ app, and body composition management program (ACCUNIQ MANAGER) |
| USB Storage | You can save or retrieve all measurement data. |

Scan the QR code on the LCD or result sheet, transmit it to the management website, and check the results.

* For the purpose of improvement, the specifications and designs of this device and options may be changed without notice.

Remote technical support with PC (ACCUNIQ REMOTE SUPPORT)

** This product is a medical diagnostic device. Read "Precautions" and "Usage Methods" carefully before use.



QR Code

Remote Support



